

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
 Data File : VX024875.D
 Acq On : 22 Oct 2021 16:36
 Operator : JC/MD
 Sample : M4291-06
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P002-TW001-03

Quant Time: Oct 23 03:55:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 23 03:44:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	72434	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	67580	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	33195	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	11586	5.448	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 109.000%		
7) Chloroethane-d5	1.672	69	10494	5.477	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 109.600%		
11) 1,1-Dichloroethene-d2	2.312	63	28625	3.611	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 72.200%		
20) 2-Butanone-d5	4.471	46	39717	51.085	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	= 102.160%		
24) Chloroform-d	5.062	84	54447	5.319	ug/L	-0.01
Spiked Amount	5.000	Range 70 - 125	Recovery	= 106.400%		
26) 1,2-Dichloroethane-d4	5.970	65	27395	4.896	ug/L	0.01
Spiked Amount	5.000	Range 70 - 130	Recovery	= 98.000%		
32) Benzene-d6	5.983	84	86232	5.476	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 109.600%		
36) 1,2-Dichloropropane-d6	7.318	67	24681	4.913	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 98.200%		
41) Toluene-d8	8.653	98	84537	5.355	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 107.000%		
43) trans-1,3-Dichloroprop...	8.958	79	10463	4.538	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 90.800%		
46) 2-Hexanone-d5	9.390	63	42473	51.181	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 102.360%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	19375	5.046	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 101.000%		
66) 1,2-Dichlorobenzene-d4	12.323	152	29968	5.357	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 107.200%		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
Data File : VX024875.D
Acq On : 22 Oct 2021 16:36
Operator : JC/MD
Sample : M4291-06
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P002-TW001-03

Quant Time: Oct 23 03:55:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 23 03:44:55 2021
Response via : Initial Calibration

