

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
 Data File : VX024877.D
 Acq On : 22 Oct 2021 17:24
 Operator : JC/MD
 Sample : M4291-08
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P009-TW001-02

Quant Time: Oct 23 03:55:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 23 03:44:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	71912	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	66755	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	31392	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	11476	5.436	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.800%
7) Chloroethane-d5	1.672	69	10809	5.682	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.600%
11) 1,1-Dichloroethene-d2	2.306	63	31186	3.963	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	79.200%
20) 2-Butanone-d5	4.477	46	38515	49.898	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.800%
24) Chloroform-d	5.068	84	53262	5.241	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.800%
26) 1,2-Dichloroethane-d4	5.964	65	28108	5.060	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.200%
32) Benzene-d6	5.983	84	86812	5.581	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.600%
36) 1,2-Dichloropropane-d6	7.318	67	24903	5.018	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.400%
41) Toluene-d8	8.653	98	82975	5.321	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.400%
43) trans-1,3-Dichloroprop...	8.958	79	10333	4.537	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	90.800%
46) 2-Hexanone-d5	9.391	63	41190	50.248	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.500%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	19072	5.029	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	100.600%
66) 1,2-Dichlorobenzene-d4	12.323	152	29087	5.498	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102221\
Data File : VX024877.D
Acq On : 22 Oct 2021 17:24
Operator : JC/MD
Sample : M4291-08
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P009-TW001-02

Quant Time: Oct 23 03:55:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 23 03:44:55 2021
Response via : Initial Calibration

