

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048288.D
 Acq On : 22 Oct 2025 11:41
 Operator : JC/MD
 Sample : Q3368-07 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PRESS-SLUDGE

Quant Time: Oct 22 12:00:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96420	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	193055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	220019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121001	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	72767	53.827	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.660%	
35) Dibromofluoromethane	5.373	113	59418	45.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.160%	
50) Toluene-d8	8.635	98	220804	47.797	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 95.600%	
62) 4-Bromofluorobenzene	11.061	95	104157	59.266	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.540%	
Target Compounds						
					Qvalue	
10) Methyl Iodide	2.453	142	29419	22.491	ug/l	99
16) Acetone	2.380	43	37840	81.592	ug/l	100
25) 2-Butanone	4.556	43	2128	3.625	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048288.D
Acq On : 22 Oct 2025 11:41
Operator : JC/MD
Sample : Q3368-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PRESS-SLUDGE

Quant Time: Oct 22 12:00:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

