

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048293.D
 Acq On : 22 Oct 2025 13:25
 Operator : JC/MD
 Sample : Q3403-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-18

Quant Time: Oct 24 05:53:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	130801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	256114	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	225488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	125813	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97488	53.158	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.320%
35) Dibromofluoromethane	5.373	113	81404	46.557	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.120%
50) Toluene-d8	8.635	98	217681	35.519	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	71.040%#
62) 4-Bromofluorobenzene	11.061	95	110780	47.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.020%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	41641	66.187	ug/l	97
43) Isopropyl Acetate	6.330	43	9321	2.755	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048293.D
Acq On : 22 Oct 2025 13:25
Operator : JC/MD
Sample : Q3403-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-18

Quant Time: Oct 24 05:53:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

