

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048294.D
 Acq On : 22 Oct 2025 13:46
 Operator : JC/MD
 Sample : Q3412-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6

Quant Time: Oct 24 05:54:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	111804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	280603	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	251721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	132683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84910	54.167	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.340%
35) Dibromofluoromethane	5.373	113	68133	35.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	71.140%#
50) Toluene-d8	8.635	98	270506	40.286	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	80.580%#
62) 4-Bromofluorobenzene	11.061	95	119813	46.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.800%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	42400	78.845	ug/l	99
20) Methylene Chloride	2.794	84	2435	1.858	ug/l #	90
43) Isopropyl Acetate	6.324	43	11374	3.068	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048294.D
Acq On : 22 Oct 2025 13:46
Operator : JC/MD
Sample : Q3412-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6

Quant Time: Oct 24 05:54:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

