

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048297.D
 Acq On : 22 Oct 2025 14:48
 Operator : JC/MD
 Sample : Q3373-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 24 05:55:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	129263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	253725	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	227885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124610	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97539	53.819	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.640%
35) Dibromofluoromethane	5.379	113	79109	45.671	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.635	98	222504	36.648	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	73.300%#
62) 4-Bromofluorobenzene	11.061	95	110652	47.907	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.820%
Target Compounds						
16) Acetone	2.380	43	21501	34.582	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048297.D
Acq On : 22 Oct 2025 14:48
Operator : JC/MD
Sample : Q3373-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 24 05:55:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

