

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048298.D
 Acq On : 22 Oct 2025 15:09
 Operator : JC/MD
 Sample : Q3373-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 24 05:55:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	120615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	206915	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	230652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	75377	44.573	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.140%
35) Dibromofluoromethane	5.373	113	76126	53.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.780%
50) Toluene-d8	8.635	98	243271	49.133	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.260%
62) 4-Bromofluorobenzene	11.061	95	121217	64.353	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	128.700%#
Target Compounds						
16) Acetone	2.380	43	27920	48.126	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048298.D
Acq On : 22 Oct 2025 15:09
Operator : JC/MD
Sample : Q3373-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 24 05:55:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

