

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048299.D
 Acq On : 22 Oct 2025 15:29
 Operator : JC/MD
 Sample : Q3373-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Oct 24 05:56:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	128102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	252845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	210146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	122933	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97398	54.228	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.460%
35) Dibromofluoromethane	5.373	113	77728	45.029	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.635	98	216127	35.721	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	71.440%#
62) 4-Bromofluorobenzene	11.061	95	108084	46.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.920%
Target Compounds						
16) Acetone	2.380	43	18154	29.463	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048299.D
Acq On : 22 Oct 2025 15:29
Operator : JC/MD
Sample : Q3373-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAGEMENT

Quant Time: Oct 24 05:56:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

