

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
 Data File : VX048300.D
 Acq On : 22 Oct 2025 15:50
 Operator : JC/MD
 Sample : Q3399-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-ROLLOFF

Quant Time: Oct 23 14:28:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	155579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	306649	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	274083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	141254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	111265	51.008	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.020%
35) Dibromofluoromethane	5.373	113	91463	43.690	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.380%
50) Toluene-d8	8.635	98	303466	41.356	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	82.720%#
62) 4-Bromofluorobenzene	11.061	95	128881	46.169	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.340%
Target Compounds						
						Qvalue
16) Acetone	2.373	43	47589	63.594	ug/l	94
20) Methylene Chloride	2.794	84	2353	1.290	ug/l #	77
43) Isopropyl Acetate	6.330	43	11664	2.879	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102225\
Data File : VX048300.D
Acq On : 22 Oct 2025 15:50
Operator : JC/MD
Sample : Q3399-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-ROLLOFF

Quant Time: Oct 23 14:28:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

