

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102318\
 Data File : VX005518.D
 Acq On : 23 Oct 2018 14:00
 Operator : JC/MD
 Sample : PB114080TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114080TB

Quant Time: Oct 24 01:43:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	227663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	360917	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131842	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	147910	58.54	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	117.08%	
35) Dibromofluoromethane	5.50	113	118498	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
50) Toluene-d8	8.72	98	377775	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
62) 4-Bromofluorobenzene	11.14	95	123437	40.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.08%	
Target Compounds						
16) Acetone	2.45	43	7578	5.187	ug/l	96
18) Methyl Acetate	2.78	43	4755	1.144	ug/l	99
25) 2-Butanone	4.70	43	6013	2.813	ug/l #	87
43) Isopropyl Acetate	6.46	43	173649	28.999	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102318\
Data File : VX005518.D
Acq On : 23 Oct 2018 14:00
Operator : JC/MD
Sample : PB114080TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114080TB

Quant Time: Oct 24 01:43:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

