

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102318\
 Data File : VX005520.D
 Acq On : 23 Oct 2018 14:54
 Operator : JC/MD
 Sample : PB114080ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB114080ZHE#02

Quant Time: Oct 24 02:06:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	225598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	370530	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	303705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126793	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	148616	59.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.72%	
35) Dibromofluoromethane	5.51	113	118748	48.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.78%	
50) Toluene-d8	8.72	98	387564	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
62) 4-Bromofluorobenzene	11.14	95	120308	38.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.02%	
Target Compounds						
16) Acetone	2.45	43	7817	5.400	ug/l #	87
18) Methyl Acetate	2.78	43	6334	1.538	ug/l	95
25) 2-Butanone	4.71	43	5442	2.569	ug/l #	83
43) Isopropyl Acetate	6.46	43	171704	27.930	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102318\
Data File : VX005520.D
Acq On : 23 Oct 2018 14:54
Operator : JC/MD
Sample : PB114080ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB114080ZHE#02

Quant Time: Oct 24 02:06:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

