

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102318\
 Data File : VX005521.D
 Acq On : 23 Oct 2018 15:20
 Operator : JC/MD
 Sample : J5622-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KG-VB-16195

Quant Time: Oct 24 02:16:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226823	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	347497	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132325	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	141475	56.20	ug/l	0.00
Spiked Amount			Recovery	=	112.40%	
35) Dibromofluoromethane	5.51	113	113578	49.35	ug/l	0.01
Spiked Amount			Recovery	=	98.70%	
50) Toluene-d8	8.72	98	377070	47.72	ug/l	0.00
Spiked Amount			Recovery	=	95.44%	
62) 4-Bromofluorobenzene	11.14	95	122130	41.15	ug/l	0.00
Spiked Amount			Recovery	=	82.30%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	32238	22.149	ug/l	93
17) Carbon Disulfide	2.57	76	6064	1.008	ug/l	96
18) Methyl Acetate	2.78	43	7763	1.875	ug/l	94
25) 2-Butanone	4.70	43	14690	6.898	ug/l	95
43) Isopropyl Acetate	6.46	43	101738	17.646	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102318\
Data File : VX005521.D
Acq On : 23 Oct 2018 15:20
Operator : JC/MD
Sample : J5622-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KG-VB-16195

Quant Time: Oct 24 02:16:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

