

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102318\
 Data File : VX005527.D
 Acq On : 23 Oct 2018 18:00
 Operator : JC/MD
 Sample : J5646-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Oct 24 03:43:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	248714	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	346262	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.12	117	282396	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	143585	52.02	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	104.04%	
35) Dibromofluoromethane	5.50	113	121222	52.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.72%	
50) Toluene-d8	8.71	98	381503	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.14	95	126100	42.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.26%	
Target Compounds						
11) Tert butyl alcohol	3.06	59	60907	78.221	ug/l	98
16) Acetone	2.45	43	3192	2.000	ug/l	90
19) Methyl tert-butyl Ether	3.20	73	30733	4.006	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102318\
Data File : VX005527.D
Acq On : 23 Oct 2018 18:00
Operator : JC/MD
Sample : J5646-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Oct 24 03:43:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

