

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102318\
 Data File : VX005515.D
 Acq On : 23 Oct 2018 12:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 24 01:29:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.391	0.389	0.5	66	0.00
3 P	Chloromethane	0.566	0.505	10.8	66	0.00
4 C	Vinyl Chloride	0.574	0.510	11.1#	68	0.00
5 T	Bromomethane	0.356	0.331	7.0	75	0.00
6 T	Chloroethane	0.342	0.340	0.6	72	0.00
7 T	Trichlorofluoromethane	0.766	0.761	0.7	74	0.00
8 T	Diethyl Ether	0.327	0.309	5.5	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.442	0.470	-6.3	78	0.00
10 T	Methyl Iodide	0.440	0.471	-7.0	69	0.00
11 T	Tert butyl alcohol	0.157	0.151	3.8	74	0.00
12 CM	1,1-Dichloroethene	0.438	0.416	5.0#	71	0.00
13 T	Acrolein	0.105	0.082	21.9#	59	0.00
14 T	Allyl chloride	0.961	0.981	-2.1	75	0.00
15 T	Acrylonitrile	0.355	0.353	0.6	79	0.00
16 T	Acetone	0.321	0.285	11.2	67	0.00
17 T	Carbon Disulfide	1.327	1.107	16.6	65	0.00
18 T	Methyl Acetate	0.913	1.013	-11.0	83	0.00
19 T	Methyl tert-butyl Ether	1.542	1.585	-2.8	76	0.00
20 T	Methylene Chloride	0.520	0.503	3.3	74	0.00
21 T	trans-1,2-Dichloroethene	0.485	0.472	2.7	77	0.00
22 T	Diisopropyl ether	1.615	1.792	-11.0	80	0.00
23 T	Vinyl Acetate	1.411	1.495	-6.0	79	0.00
24 P	1,1-Dichloroethane	0.926	0.948	-2.4	80	0.00
25 T	2-Butanone	0.469	0.483	-3.0	74	-0.01
26 T	2,2-Dichloropropane	0.692	0.778	-12.4	88	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.542	-2.8	80	0.00
28 T	Bromochloromethane	0.422	0.475	-12.6	84	0.00
29 T	Tetrahydrofuran	0.299	0.330	-10.4	77	0.00
30 C	Chloroform	0.859	0.930	-8.3#	81	0.00
31 T	Cyclohexane	0.763	0.795	-4.2	74	0.00
32 T	1,1,1-Trichloroethane	0.747	0.775	-3.7	78	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.594	-7.0	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.331	0.355	-7.3	83	0.00
36 T	1,1-Dichloropropene	0.484	0.475	1.9	75	0.00
37 T	Ethyl Acetate	0.623	0.647	-3.9	80	-0.01
38 T	Carbon Tetrachloride	0.487	0.490	-0.6	78	0.00
39 T	Methylcyclohexane	0.524	0.498	5.0	72	0.00
40 TM	Benzene	1.453	1.362	6.3	69	0.00
41 T	Methacrylonitrile	0.334	0.349	-4.5	79	0.00
42 TM	1,2-Dichloroethane	0.513	0.519	-1.2	78	0.00
43 T	Isopropyl Acetate	0.830	0.944	-13.7	85	-0.01
44 TM	Trichloroethene	0.416	0.354	14.9	73	0.00
45 C	1,2-Dichloropropane	0.356	0.367	-3.1#	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102318\
 Data File : VX005515.D
 Acq On : 23 Oct 2018 12:13
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 24 01:29:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.226	3.8	73	0.00
47 T	Bromodichloromethane	0.435	0.454	-4.4	80	0.00
48 T	Methyl methacrylate	0.430	0.449	-4.4	76	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	64	0.00
50 S	Toluene-d8	1.137	1.192	-4.8	80	0.00
51 T	4-Methyl-2-Pentanone	0.574	0.614	-7.0	80	0.00
52 CM	Toluene	0.795	0.803	-1.0#	75	0.00
53 T	t-1,3-Dichloropropene	0.471	0.508	-7.9	79	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.544	-6.7	79	0.00
55 T	1,1,2-Trichloroethane	0.337	0.346	-2.7	80	0.00
56 T	Ethyl methacrylate	0.491	0.538	-9.6	79	0.00
57 T	1,3-Dichloropropane	0.562	0.576	-2.5	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.275	0.305	-10.9	80	0.00
59 T	2-Hexanone	0.462	0.483	-4.5	79	0.00
60 T	Dibromochloromethane	0.348	0.364	-4.6	78	0.00
61 T	1,2-Dibromoethane	0.353	0.355	-0.6	77	0.00
62 S	4-Bromofluorobenzene	0.427	0.457	-7.0	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	75	0.00
64 T	Tetrachloroethene	0.403	0.412	-2.2	79	0.00
65 PM	Chlorobenzene	1.064	1.042	2.1	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.393	-3.7	78	0.00
67 C	Ethyl Benzene	1.729	1.801	-4.2#	77	0.00
68 T	m/p-Xylenes	0.671	0.695	-3.6	76	0.00
69 T	o-Xylene	0.647	0.668	-3.2	76	0.00
70 T	Styrene	1.065	1.134	-6.5	77	0.00
71 P	Bromoform	0.339	0.346	-2.1	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	2.972	3.195	-7.5	79	0.00
74 T	N-amyl acetate	1.434	1.549	-8.0	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.121	1.100	1.9	79	0.00
76 T	1,2,3-Trichloropropane	0.970	1.023	-5.5	82	0.00
77 T	Bromobenzene	0.850	0.841	1.1	77	0.00
78 T	n-propylbenzene	3.420	3.702	-8.2	80	0.00
79 T	2-Chlorotoluene	2.095	2.184	-4.2	79	0.00
80 T	1,3,5-Trimethylbenzene	2.545	2.741	-7.7	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.319	0.331	-3.8	79	0.00
82 T	4-Chlorotoluene	2.483	2.610	-5.1	79	0.00
83 T	tert-Butylbenzene	2.578	2.691	-4.4	78	0.00
84 T	1,2,4-Trimethylbenzene	2.619	2.834	-8.2	79	0.00
85 T	sec-Butylbenzene	3.050	3.292	-7.9	79	0.00
86 T	p-Isopropyltoluene	2.757	2.979	-8.1	78	0.00
87 T	1,3-Dichlorobenzene	1.588	1.579	0.6	78	0.00
88 T	1,4-Dichlorobenzene	1.625	1.613	0.7	79	0.00
89 T	n-Butylbenzene	2.503	2.693	-7.6	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102318\
Data File : VX005515.D
Acq On : 23 Oct 2018 12:13
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 24 01:29:29 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.485	-2.1	79	0.00
91 T	1,2-Dichlorobenzene	1.621	1.607	0.9	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.274	0.268	2.2	78	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.235	-3.9	78	0.00
94 T	Hexachlorobutadiene	0.628	0.622	1.0	79	0.00
95 T	Naphthalene	3.448	3.735	-8.3	77	0.00
96 T	1,2,3-Trichlorobenzene	1.217	1.247	-2.5	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6