

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102318\
 Data File : VX005516.D
 Acq On : 23 Oct 2018 12:58
 Operator : JC/MD
 Sample : VX1023WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1023WBL01

Quant Time: Oct 24 01:35:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	195898	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282959	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	243189	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	115172	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	124689	57.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.70%	
35) Dibromofluoromethane	5.50	113	95795	51.11	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.22%	
50) Toluene-d8	8.71	98	325411	50.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.16%	
62) 4-Bromofluorobenzene	11.14	95	105297	43.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.14%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102318\
Data File : VX005516.D
Acq On : 23 Oct 2018 12:58
Operator : JC/MD
Sample : VX1023WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1023WBL01

Quant Time: Oct 24 01:35:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX005516.D

Acq: 23 Oct 2018 12:58

Instrument :

MSVOA_X

ClientSampleId :


VX1023WBL01

Tot Ion:168 Resp: 195898

Ion Ratio Lower Upper

168 100

99 49.7 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 57.350 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.01 min

Lab File: VX005516.D


Acq: 23 Oct 2018 12:58

Tgt Ion: 65 Resp: 124689

Ion Ratio Lower Upper

65 100

67 51.7 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

60000

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.86 min Scan# 936
Delta R.T. -0.01 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58

Instrument :
MSVOA_X
ClientSampleId :
VX1023WBL01

Tot Ion:114 Resp: 282959
Ion Ratio Lower Upper
114 100
63 20.3 0.0 39.0
88 15.3 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VX0
Ion 87.90 (87.60 to 88.60): VX0






#35
Dibromofluoromethane
Concen: 51.115 ug/l
RT: 5.50 min Scan# 713
Delta R.T. 0.00 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58

Tgt Ion:113 Resp: 95795
Ion Ratio Lower Upper
113 100
111 104.7 82.4 123.6
192 24.1 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.577 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. 0.00 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58




Instrument :
MSVOA_X
ClientSampleId :
VX1023WBL01


Tot Ion: 98 Resp: 325411
Ion Ratio Lower Upper
98 100
100 62.8 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 43.565 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58

Tgt Ion: 95 Resp: 105297
Ion Ratio Lower Upper
95 100
174 92.8 0.0 185.2
176 88.8 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. 0.00 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58

Instrument :
MSVOA_X
ClientSampleId :
VX1023WBL01

Tot Ion	Ratio	Lower	Upper
117	100		
82	53.9	47.8	71.6
119	31.5	29.3	43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. 0.00 min
Lab File: VX005516.D
Acq: 23 Oct 2018 12:58

Tgt Ion	Ratio	Lower	Upper
152	100		
115	55.4	39.0	117.0
150	157.6	0.0	351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

