


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102318\
Data File : VX005514.D
Acq On : 23 Oct 2018 11:41
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	3637	PASS
75	95	30	60	50.9	9639	PASS
95	95	100	100	100.0	18946	PASS
96	95	5	9	7.2	1356	PASS
173	174	0.00	2	1.6	286	PASS
174	95	50	100	96.4	18265	PASS
175	174	5	9	7.6	1392	PASS
176	174	95	101	95.2	17386	PASS
177	176	5	9	7.2	1248	PASS