

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102320\
 Data File : VX019081.D
 Acq On : 23 Oct 2020 17:16
 Operator : JC/SP
 Sample : L4473-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT-TB-20201019

Quant Time: Oct 26 04:37:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X102320W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Oct 26 03:21:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	4.98	128	38958	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.83	114	220036	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.10	117	193642	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	6.03	65	81281	31.32	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	104.40%
60) 4-Bromofluorobenzene	11.12	95	85859	29.54	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	98.47%
63) Toluene-d8	8.70	98	259379	30.34	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.13%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102320\
Data File : VX019081.D
Acq On : 23 Oct 2020 17:16
Operator : JC/SP
Sample : L4473-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT-TB-20201019

Quant Time: Oct 26 04:37:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X102320W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Oct 26 03:21:29 2020
Response via : Initial Calibration

