

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102320\
 Data File : VX019097.D
 Acq On : 24 Oct 2020 01:30
 Operator : JC/SP
 Sample : L4417-13DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-14DL

Quant Time: Oct 29 08:32:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	268861	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	465148	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438625	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	226637	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	163348	50.72	ug/l	0.00
Spiked Amount			Recovery	=	101.44%	
35) Dibromofluoromethane	5.46	113	143366	48.68	ug/l	-0.01
Spiked Amount			Recovery	=	97.36%	
50) Toluene-d8	8.70	98	572042	50.90	ug/l	0.00
Spiked Amount			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.12	95	203678	50.95	ug/l	0.00
Spiked Amount			Recovery	=	101.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.02	59	67188	103.008	ug/l	99
16) Acetone	2.43	43	2176	1.984	ug/l	98
19) Methyl tert-butyl Ether	3.17	73	751013	90.599	ug/l	100
40) Benzene	6.11	78	105245	8.562	ug/l	99
52) Toluene	8.76	92	4270	0.526	ug/l	100
67) Ethyl Benzene	10.24	91	36580	2.307	ug/l	96
68) m/p-Xylenes	10.34	106	1818	0.299	ug/l	87
73) Isopropylbenzene	11.00	105	13963	0.856	ug/l	98
78) n-propylbenzene	11.34	91	10821	0.586	ug/l	96
84) 1,2,4-Trimethylbenzene	11.79	105	3006	0.223	ug/l	100
95) Naphthalene	13.82	128	11553	0.713	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102320\
Data File : VX019097.D
Acq On : 24 Oct 2020 01:30
Operator : JC/SP
Sample : L4417-13DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-14DL

Quant Time: Oct 29 08:32:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

