

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102320\
 Data File : VX019098.D
 Acq On : 24 Oct 2020 01:52
 Operator : JC/SP
 Sample : L4417-16DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5DL

Quant Time: Oct 26 06:06:30 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	273785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	475315	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	448914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231973	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	166241	50.69	ug/l	0.00
Spiked Amount			Recovery	=	101.38%	
35) Dibromofluoromethane	5.46	113	144225	47.93	ug/l	0.00
Spiked Amount			Recovery	=	95.86%	
50) Toluene-d8	8.70	98	584575	50.90	ug/l	0.00
Spiked Amount			Recovery	=	101.80%	
62) 4-Bromofluorobenzene	11.12	95	211552	51.79	ug/l	0.00
Spiked Amount			Recovery	=	103.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.01	59	105700	159.137	ug/l	99
16) Acetone	2.42	43	4753	4.255	ug/l	93
19) Methyl tert-butyl Ether	3.17	73	97804	11.587	ug/l #	72
31) Cyclohexane	5.55	56	30164	7.854	ug/l #	95
39) Methylcyclohexane	7.44	83	18221	3.813	ug/l	98
40) Benzene	6.11	78	1117609	88.975	ug/l	100
52) Toluene	8.76	92	61938	7.466	ug/l	99
67) Ethyl Benzene	10.23	91	1354928	83.476	ug/l	100
68) m/p-Xylenes	10.34	106	23698	3.811	ug/l	100
73) Isopropylbenzene	11.00	105	302739	18.142	ug/l	100
78) n-propylbenzene	11.34	91	307729	16.274	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	16369	1.188	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	3153	0.228	ug/l	100
85) sec-Butylbenzene	11.93	105	11374	0.692	ug/l	98
89) n-Butylbenzene	12.37	91	16630	1.210	ug/l #	76
95) Naphthalene	13.82	128	391274	23.581	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102320\
Data File : VX019098.D
Acq On : 24 Oct 2020 01:52
Operator : JC/SP
Sample : L4417-16DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5DL

Quant Time: Oct 26 06:06:30 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

