

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102320\
 Data File : VX019106.D
 Acq On : 24 Oct 2020 04:51
 Operator : JC/SP
 Sample : L4470-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-MW179D-GW-20201015

Quant Time: Nov 03 11:03:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	262370	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	453077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	428820	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	214987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	159877	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
35) Dibromofluoromethane	5.46	113	138768	48.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.76%	
50) Toluene-d8	8.70	98	558407	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	11.12	95	196361	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.36	96	1069	0.419	ug/l	96
16) Acetone	2.42	43	2424	2.264	ug/l	98
19) Methyl tert-butyl Ether	3.17	73	6755	0.835	ug/l	95
21) trans-1,2-Dichloroethene	3.14	96	1221	0.410	ug/l #	81
27) cis-1,2-Dichloroethene	4.56	96	115513	34.157	ug/l	90
44) Trichloroethene	7.19	130	456909	130.855	ug/l	99
64) Tetrachloroethene	9.32	164	53774	16.470	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102320\
Data File : VX019106.D
Acq On : 24 Oct 2020 04:51
Operator : JC/SP
Sample : L4470-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-MW179D-GW-20201015

Quant Time: Nov 03 11:03:06 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

