

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102320\
 Data File : VX019109.D
 Acq On : 24 Oct 2020 05:58
 Operator : JC/SP
 Sample : PB132563ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB132563ZHE#08

Quant Time: Oct 26 06:31:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 22 17:51:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	268574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	469060	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	447328	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	225741	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	164931	51.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.54%	
35) Dibromofluoromethane	5.46	113	145545	49.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.02%	
50) Toluene-d8	8.70	98	577376	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	11.12	95	204157	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
Target Compounds						
16) Acetone	2.42	43	9959	9.089	ug/l	96
20) Methylene Chloride	2.84	84	4543	1.489	ug/l	91
43) Isopropyl Acetate	6.41	43	60900	10.010	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102320\
Data File : VX019109.D
Acq On : 24 Oct 2020 05:58
Operator : JC/SP
Sample : PB132563ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB132563ZHE#08

Quant Time: Oct 26 06:31:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102120W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 22 17:51:42 2020
Response via : Initial Calibration

