

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
 Data File : VX024895.D
 Acq On : 23 Oct 2021 15:01
 Operator : JC/MD
 Sample : M4303-01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P003-TW001-02

Quant Time: Oct 25 02:33:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 25 02:31:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	84893	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	77391	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	36692	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	11552	4.635	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.800%
7) Chloroethane-d5	1.672	69	11907	5.303	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.000%
11) 1,1-Dichloroethene-d2	2.306	63	30879	3.324	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.400%
20) 2-Butanone-d5	4.477	46	49143	53.932	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.860%
24) Chloroform-d	5.062	84	55664	4.639	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.800%
26) 1,2-Dichloroethane-d4	5.970	65	29793	4.543	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
32) Benzene-d6	5.982	84	87381	4.846	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
36) 1,2-Dichloropropane-d6	7.318	67	27867	4.844	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.800%
41) Toluene-d8	8.653	98	81118	4.487	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
43) trans-1,3-Dichloroprop...	8.958	79	10052	3.807	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	9.390	63	46429	48.855	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.720%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	20687	4.705	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	27094	4.382	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
Data File : VX024895.D
Acq On : 23 Oct 2021 15:01
Operator : JC/MD
Sample : M4303-01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P003-TW001-02

Quant Time: Oct 25 02:33:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 25 02:31:08 2021
Response via : Initial Calibration

