

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
 Data File : VX024896.D
 Acq On : 23 Oct 2021 15:25
 Operator : JC/MD
 Sample : M4303-02
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P004-TW001-02

Quant Time: Oct 25 02:33:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 25 02:31:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	78051	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	72141	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	34587	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	12363	5.395	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.000%
7) Chloroethane-d5	1.672	69	12321	5.968	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.400%
11) 1,1-Dichloroethene-d2	2.313	63	31809	3.724	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.400%
20) 2-Butanone-d5	4.477	46	50485	60.262	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	120.520%
24) Chloroform-d	5.068	84	56500	5.122	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
26) 1,2-Dichloroethane-d4	5.964	65	30437	5.048	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.000%
32) Benzene-d6	5.983	84	88767	5.281	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.600%
36) 1,2-Dichloropropane-d6	7.318	67	29123	5.431	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	108.600%
41) Toluene-d8	8.653	98	83580	4.959	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.200%
43) trans-1,3-Dichloroprop...	8.958	79	10952	4.449	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	9.391	63	48033	54.221	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	108.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	21157	5.162	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	103.200%
66) 1,2-Dichlorobenzene-d4	12.323	152	29207	5.011	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
Data File : VX024896.D
Acq On : 23 Oct 2021 15:25
Operator : JC/MD
Sample : M4303-02
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P004-TW001-02

Quant Time: Oct 25 02:33:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 25 02:31:08 2021
Response via : Initial Calibration

