

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
 Data File : VX024900.D
 Acq On : 23 Oct 2021 17:00
 Operator : JC/MD
 Sample : M4303-06
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P012-TW001-02

Quant Time: Oct 25 02:34:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 25 02:31:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	82628	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.061	117	74291	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	34989	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	11027	4.546	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	91.000%
7) Chloroethane-d5	1.672	69	11815	5.406	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.200%
11) 1,1-Dichloroethene-d2	2.313	63	31345	3.466	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.400%
20) 2-Butanone-d5	4.483	46	50115	56.506	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.020%
24) Chloroform-d	5.068	84	56622	4.849	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.000%
26) 1,2-Dichloroethane-d4	5.964	65	31184	4.886	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	5.989	84	91296	5.274	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.400%
36) 1,2-Dichloropropane-d6	7.312	67	27786	5.031	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.600%
41) Toluene-d8	8.653	98	83652	4.820	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.400%
43) trans-1,3-Dichloroprop...	8.958	79	10234	4.037	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	9.390	63	46274	50.724	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	101.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	21029	4.982	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.600%
66) 1,2-Dichlorobenzene-d4	12.323	152	28388	4.814	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.200%
Target Compounds						Qvalue
3) Chloromethane	1.288	50	1051	0.317	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
Data File : VX024900.D
Acq On : 23 Oct 2021 17:00
Operator : JC/MD
Sample : M4303-06
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P012-TW001-02

Quant Time: Oct 25 02:34:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 25 02:31:08 2021
Response via : Initial Calibration

