

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
 Data File : VX024906.D
 Acq On : 23 Oct 2021 19:23
 Operator : JC/MD
 Sample : M4303-14
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P019-TW001-02

Quant Time: Oct 25 02:35:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Oct 25 02:31:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	81552	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	73786	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	33596	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	10851	4.532	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.600%
7) Chloroethane-d5	1.672	69	10326	4.787	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.800%
11) 1,1-Dichloroethene-d2	2.306	63	27455	3.076	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.600%
20) 2-Butanone-d5	4.483	46	47597	54.375	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.760%
24) Chloroform-d	5.062	84	53605	4.651	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.000%
26) 1,2-Dichloroethane-d4	5.970	65	27568	4.376	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.600%
32) Benzene-d6	5.983	84	84167	4.896	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.000%
36) 1,2-Dichloropropane-d6	7.318	67	26159	4.769	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	8.653	98	76678	4.448	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.000%
43) trans-1,3-Dichloroprop...	8.951	79	9543	3.791	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.800%
46) 2-Hexanone-d5	9.390	63	45504	50.221	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.440%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	20427	4.873	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	12.323	152	27181	4.801	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.000%
Target Compounds					Qvalue	
13) Acetone	2.410	43	1878	2.860	ug/L	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102321\
Data File : VX024906.D
Acq On : 23 Oct 2021 19:23
Operator : JC/MD
Sample : M4303-14
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P019-TW001-02

Quant Time: Oct 25 02:35:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Oct 25 02:31:08 2021
Response via : Initial Calibration

