

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038429.D
 Acq On : 23 Oct 2023 13:55
 Operator : JC/MD
 Sample : 05007-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-22-37-40

Quant Time: Oct 24 00:01:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134390	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	240375	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129547	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95080	43.986	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.980%
35) Dibromofluoromethane	5.385	113	80604	46.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.080%
50) Toluene-d8	8.647	98	297938	45.015	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.040%#
62) 4-Bromofluorobenzene	11.079	95	114620	45.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.360%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	2841	3.149	ug/l	98
30) Chloroform	5.093	83	82492	26.968	ug/l	98
44) Trichloroethene	7.135	130	4861	2.782	ug/l	83
64) Tetrachloroethene	9.275	164	4661	2.934	ug/l	92
91) 1,2-Dichlorobenzene	12.335	146	3249	0.793	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038429.D
Acq On : 23 Oct 2023 13:55
Operator : JC/MD
Sample : 05007-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-22-37-40

Quant Time: Oct 24 00:01:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

