

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038432.D
 Acq On : 23 Oct 2023 15:04
 Operator : JC/MD
 Sample : 05007-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-23-37-40

Quant Time: Oct 24 00:02:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126400	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228120	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121030	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90874	44.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.400%
35) Dibromofluoromethane	5.385	113	75763	46.092	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.180%
50) Toluene-d8	8.653	98	283207	45.088	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.180%#
62) 4-Bromofluorobenzene	11.079	95	107385	45.096	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.200%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	2378	2.803	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	47049	10.215	ug/l	95
30) Chloroform	5.098	83	1866	0.649	ug/l	98
44) Trichloroethene	7.128	130	6380	3.848	ug/l	90
64) Tetrachloroethene	9.274	164	12044	7.964	ug/l	93
65) Chlorobenzene	10.079	112	1849	0.378	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038432.D
Acq On : 23 Oct 2023 15:04
Operator : JC/MD
Sample : 05007-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-23-37-40

Quant Time: Oct 24 00:02:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

