

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038433.D
 Acq On : 23 Oct 2023 15:28
 Operator : JC/MD
 Sample : 05007-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-101923

Quant Time: Oct 24 00:03:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128204	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	234329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237602	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123571	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92194	44.709	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.420%
35) Dibromofluoromethane	5.391	113	76955	45.576	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.160%
50) Toluene-d8	8.647	98	290845	45.077	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.160%#
62) 4-Bromofluorobenzene	11.079	95	107857	44.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038433.D
Acq On : 23 Oct 2023 15:28
Operator : JC/MD
Sample : 05007-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-101923

Quant Time: Oct 24 00:03:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

