

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038434.D
 Acq On : 23 Oct 2023 15:51
 Operator : JC/MD
 Sample : 05007-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-24-23-26

Quant Time: Oct 24 00:03:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	127479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231600	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94002	45.845	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.700%
35) Dibromofluoromethane	5.391	113	77392	46.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.760%
50) Toluene-d8	8.647	98	289218	45.353	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.700%#
62) 4-Bromofluorobenzene	11.079	95	108241	44.772	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.540%
Target Compounds						
					Qvalue	
30) Chloroform	5.105	83	2443	0.842	ug/l #	83
44) Trichloroethene	7.129	130	2334	1.386	ug/l	76
64) Tetrachloroethene	9.275	164	5079	3.329	ug/l	97
93) 1,2,4-Trichlorobenzene	13.591	180	1758	0.682	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038434.D
Acq On : 23 Oct 2023 15:51
Operator : JC/MD
Sample : 05007-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-24-23-26

Quant Time: Oct 24 00:03:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

