

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038437.D
 Acq On : 23 Oct 2023 17:01
 Operator : JC/MD
 Sample : 05007-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 24 00:04:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227480	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	122009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88923	43.905	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.820%
35) Dibromofluoromethane	5.385	113	76587	46.724	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.440%
50) Toluene-d8	8.647	98	283428	45.250	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.500%#
62) 4-Bromofluorobenzene	11.079	95	107659	45.338	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.680%
Target Compounds						
					Qvalue	
3) Chloromethane	1.300	50	1070	0.864	ug/l	91
93) 1,2,4-Trichlorobenzene	13.585	180	18518	7.159	ug/l	95
96) 1,2,3-Trichlorobenzene	13.957	180	3908	1.560	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038437.D
 Acq On : 23 Oct 2023 17:01
 Operator : JC/MD
 Sample : 05007-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 24 00:04:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

