

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038439.D
 Acq On : 23 Oct 2023 17:47
 Operator : JC/MD
 Sample : 05007-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-25-23-26

Quant Time: Oct 24 00:05:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129025	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233594	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91730	44.201	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.400%
35) Dibromofluoromethane	5.385	113	77694	46.159	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	290713	45.199	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.400%#
62) 4-Bromofluorobenzene	11.079	95	111078	45.554	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.100%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.117	73	3479	0.740	ug/l	91
30) Chloroform	5.105	83	3929	1.338	ug/l	92
44) Trichloroethene	7.129	130	2650	1.561	ug/l	94
64) Tetrachloroethene	9.275	164	17263	11.112	ug/l	97
65) Chlorobenzene	10.079	112	124189	24.730	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	38820	9.205	ug/l	97
88) 1,4-Dichlorobenzene	12.042	146	259337	60.234	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	94062	23.204	ug/l	97
93) 1,2,4-Trichlorobenzene	13.597	180	7227278	2658.146	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	1606394	610.185	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038439.D
Acq On : 23 Oct 2023 17:47
Operator : JC/MD
Sample : 05007-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-25-23-26

Quant Time: Oct 24 00:05:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

