

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038442.D
 Acq On : 23 Oct 2023 18:56
 Operator : JC/MD
 Sample : 05007-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-26-10-13

Quant Time: Oct 24 00:06:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	231762	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	125347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91898	44.453	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	88.900%		
35) Dibromofluoromethane	5.391	113	76747	45.957	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.920%		
50) Toluene-d8	8.647	98	288010	45.132	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	90.260%#		
62) 4-Bromofluorobenzene	11.079	95	109962	45.452	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.900%		
Target Compounds						
44) Trichloroethene	7.135	130	2794	1.658	ug/l	86
64) Tetrachloroethene	9.275	164	1333	0.877	ug/l #	80
93) 1,2,4-Trichlorobenzene	13.591	180	22292	8.388	ug/l	96
96) 1,2,3-Trichlorobenzene	13.963	180	5323	2.069	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038442.D
Acq On : 23 Oct 2023 18:56
Operator : JC/MD
Sample : 05007-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-26-10-13

Quant Time: Oct 24 00:06:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

