

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
 Data File : VX038445.D
 Acq On : 23 Oct 2023 20:06
 Operator : JC/MD
 Sample : 04980-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BRIDAL-SHOP-BAY-SUMP

Quant Time: Oct 24 00:08:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131652	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	236898	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	123010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93753	44.275	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.540%
35) Dibromofluoromethane	5.385	113	77989	45.688	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	8.647	98	291789	44.733	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.460%#
62) 4-Bromofluorobenzene	11.079	95	110424	44.654	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.300%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.501	96	1177	0.663	ug/l	89
64) Tetrachloroethene	9.275	164	139118	88.713	ug/l	99
93) 1,2,4-Trichlorobenzene	13.591	180	8712	3.340	ug/l	96
96) 1,2,3-Trichlorobenzene	13.963	180	2207	0.874	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102323\
Data File : VX038445.D
Acq On : 23 Oct 2023 20:06
Operator : JC/MD
Sample : 04980-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BRIDAL-SHOP-BAY-SUMP

Quant Time: Oct 24 00:08:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

