

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102325\
 Data File : VX048311.D
 Acq On : 23 Oct 2025 09:46
 Operator : JC/MD
 Sample : VX1023MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1023MBL01

Quant Time: Oct 24 05:38:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	86084	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	174731	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	182476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.000	152	89788	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	75969	62.942	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	125.880%#
35) Dibromofluoromethane	5.367	113	53778	45.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.160%
50) Toluene-d8	8.628	98	220541	52.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.500%
62) 4-Bromofluorobenzene	11.061	95	90969	57.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102325\
Data File : VX048311.D
Acq On : 23 Oct 2025 09:46
Operator : JC/MD
Sample : VX1023MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1023MBL01

Quant Time: Oct 24 05:38:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

