

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102325\
 Data File : VX048335.D
 Acq On : 23 Oct 2025 18:32
 Operator : JC/MD
 Sample : Q3388-01
 Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 461

Quant Time: Oct 24 05:51:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.531	168	85331	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.744	114	173055	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	180555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	94013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	73930	61.794	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	123.580%#
35) Dibromofluoromethane	5.373	113	50423	42.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.360%
50) Toluene-d8	8.634	98	214604	51.823	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.640%
62) 4-Bromofluorobenzene	11.067	95	94671	60.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	120.180%
Target Compounds						
17) Carbon Disulfide	2.507	76	3919	1.388	ug/l	Qvalue 97
20) Methylene Chloride	2.788	84	1466	1.466	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102325\
Data File : VX048335.D
Acq On : 23 Oct 2025 18:32
Operator : JC/MD
Sample : Q3388-01
Misc : 1.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
461

Quant Time: Oct 24 05:51:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

