

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102418\
 Data File : VX005544.D
 Acq On : 24 Oct 2018 16:33
 Operator : JC/MD
 Sample : J5649-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-01

Quant Time: Nov 05 00:08:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	214170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	323283	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	287404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	137376	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	138007	52.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.58%	
35) Dibromofluoromethane	5.50	113	108519	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
50) Toluene-d8	8.71	98	384555	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
62) 4-Bromofluorobenzene	11.14	95	127857	43.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.08%	
Target Compounds						
16) Acetone	2.45	43	5380	3.894	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX102418\
Data File : VX005544.D
Acq On : 24 Oct 2018 16:33
Operator : JC/MD
Sample : J5649-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-01

Quant Time: Nov 05 00:08:37 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

