

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005550.D
 Acq On : 24 Oct 2018 19:12
 Operator : JC/MD
 Sample : J5655-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0007

Quant Time: Oct 25 05:33:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	206139	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	275674	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	128610	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	130520	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
35) Dibromofluoromethane	5.50	113	103614	50.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.86%	
50) Toluene-d8	8.71	98	365906	49.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	11.14	95	119560	42.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.78%	
Target Compounds						
16) Acetone	2.45	43	7218	5.428	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005550.D
Acq On : 24 Oct 2018 19:12
Operator : JC/MD
Sample : J5655-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0007

Quant Time: Oct 25 05:33:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

