

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005564.D
 Acq On : 25 Oct 2018 02:18
 Operator : JC/MD
 Sample : J5198-28
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-102318-D

Quant Time: Oct 25 06:44:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	209781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	311897	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	288122	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133810	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	132200	51.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.24%	
35) Dibromofluoromethane	5.51	113	105597	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
50) Toluene-d8	8.71	98	379234	49.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.94%	
62) 4-Bromofluorobenzene	11.14	95	125379	43.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.50%	
Target Compounds						
16) Acetone	2.44	43	49811	36.811	ug/l	96
20) Methylene Chloride	2.85	84	26923	11.880	ug/l	91
43) Isopropyl Acetate	6.46	43	88697	15.324	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005564.D
Acq On : 25 Oct 2018 02:18
Operator : JC/MD
Sample : J5198-28
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-102318-D

Quant Time: Oct 25 06:44:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

