

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005565.D
 Acq On : 25 Oct 2018 02:44
 Operator : JC/MD
 Sample : J5198-30
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-01-102318-E

Quant Time: Oct 25 06:45:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	196669	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	303895	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	279896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131529	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128749	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
35) Dibromofluoromethane	5.50	113	102255	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
50) Toluene-d8	8.71	98	370086	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
62) 4-Bromofluorobenzene	11.14	95	121450	43.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.00%	
Target Compounds						
16) Acetone	2.44	43	10721	8.451	ug/l	89
20) Methylene Chloride	2.86	84	20154	9.486	ug/l	94
43) Isopropyl Acetate	6.46	43	135677	24.058	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005565.D
Acq On : 25 Oct 2018 02:44
Operator : JC/MD
Sample : J5198-30
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-01-102318-E

Quant Time: Oct 25 06:45:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

