

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005567.D
 Acq On : 25 Oct 2018 03:38
 Operator : JC/MD
 Sample : J5200-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-02-102318-B

Quant Time: Oct 25 06:53:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	199267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	306690	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	284356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	131860	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
35) Dibromofluoromethane	5.50	113	103431	49.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.62%	
50) Toluene-d8	8.72	98	373683	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	11.14	95	124392	44.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.28%	
Target Compounds						
16) Acetone	2.44	43	240367	187.008	ug/l	95
18) Methyl Acetate	2.78	43	8415	1.224	ug/l	94
20) Methylene Chloride	2.86	84	19997	9.289	ug/l	90
43) Isopropyl Acetate	6.46	43	127930	22.478	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005567.D
Acq On : 25 Oct 2018 03:38
Operator : JC/MD
Sample : J5200-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-02-102318-B

Quant Time: Oct 25 06:53:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

