

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005571.D
 Acq On : 25 Oct 2018 05:24
 Operator : JC/MD
 Sample : J5655-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0004

Quant Time: Oct 25 08:21:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	298724	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	273797	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	131583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	129078	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
35) Dibromofluoromethane	5.50	113	106108	52.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.92%	
50) Toluene-d8	8.72	98	365880	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.14	95	122043	44.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.92%	
Target Compounds						
16) Acetone	2.44	43	7012	5.362	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005571.D
Acq On : 25 Oct 2018 05:24
Operator : JC/MD
Sample : J5655-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0004

Quant Time: Oct 25 08:21:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

