

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005575.D
 Acq On : 25 Oct 2018 07:10
 Operator : JC/MD
 Sample : J4771-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-08-102318

Manual Integrations
 APPROVED

MMDadoda
 10/26/2018 11:54:05 AM

Quant Time: Oct 25 08:30:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 24 13:13:34 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	205405	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	317036	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	293156	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138275	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	131929	52.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.24%	
35) Dibromofluoromethane	5.50	113	103524	48.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.44%	
50) Toluene-d8	8.71	98	388015	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	11.14	95	128114	43.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.96%	
Target Compounds						
16) Acetone	2.45	43	8977	6.775	ug/l	96
20) Methylene Chloride	2.84	84	24045099m	10835.924	ug/l	
43) Isopropyl Acetate	6.46	43	139557	23.721	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005575.D
Acq On : 25 Oct 2018 07:10
Operator : JC/MD
Sample : J4771-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-08-102318

Manual Integrations
APPROVED

MMDadoda
10/26/2018 11:54:05 AM

Quant Time: Oct 25 08:30:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 24 13:13:34 2018
Response via : Initial Calibration

