

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102418\
 Data File : VX005578.D
 Acq On : 25 Oct 2018 08:30
 Operator : JC/MD
 Sample : J5639-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BS-SOIL-3

Quant Time: Oct 26 13:24:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	308749	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	285838	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133640	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	135632	54.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.76%	
35) Dibromofluoromethane	5.50	113	104331	49.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.82%	
50) Toluene-d8	8.71	98	375022	49.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.84%	
62) 4-Bromofluorobenzene	11.14	95	123668	43.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.18%	
Target Compounds						
16) Acetone	2.44	43	9071	6.946	ug/l	99
18) Methyl Acetate	2.78	43	8589	1.237	ug/l	100
20) Methylene Chloride	2.86	84	29266	13.380	ug/l	96
25) 2-Butanone	4.70	43	1834	0.922	ug/l #	74
43) Isopropyl Acetate	6.46	43	139535	24.354	ug/l	99
95) Naphthalene	13.83	128	6486	0.711	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102418\
Data File : VX005578.D
Acq On : 25 Oct 2018 08:30
Operator : JC/MD
Sample : J5639-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BS-SOIL-3

Quant Time: Oct 26 13:24:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

