

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102419\
 Data File : VX013184.D
 Acq On : 23 Oct 2019 12:24
 Operator : JC/SP
 Sample : VX1024WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1024WBL01

Quant Time: Oct 24 04:10:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	170135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	262083	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	231119	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103626	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	106016	29.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	59.58%	
35) Dibromofluoromethane	5.48	113	83221	28.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	56.94%	
50) Toluene-d8	8.71	98	313065	28.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	57.80%	
62) 4-Bromofluorobenzene	11.13	95	114232	26.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	53.30%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102419\
Data File : VX013184.D
Acq On : 23 Oct 2019 12:24
Operator : JC/SP
Sample : VX1024WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1024WBL01

Quant Time: Oct 24 04:10:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

