

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102419\
 Data File : VX013192.D
 Acq On : 23 Oct 2019 17:22
 Operator : JC/SP
 Sample : VX1024WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1024WBL02

Quant Time: Oct 24 04:14:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	86964	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	130846	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	128479	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	78325	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	100443	55.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.44%	
35) Dibromofluoromethane	5.49	113	76776	52.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.24%	
50) Toluene-d8	8.71	98	282847	52.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.62%	
62) 4-Bromofluorobenzene	11.13	95	103743	48.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.94%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102419\
Data File : VX013192.D
Acq On : 23 Oct 2019 17:22
Operator : JC/SP
Sample : VX1024WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1024WBL02

Quant Time: Oct 24 04:14:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

