

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102419\
 Data File : VX013196.D
 Acq On : 23 Oct 2019 18:55
 Operator : JC/SP
 Sample : K5451-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW0542

Quant Time: Oct 24 04:22:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	83109	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	124235	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	120485	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	74938	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	93387	53.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.44%	
35) Dibromofluoromethane	5.49	113	71068	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
50) Toluene-d8	8.71	98	261834	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
62) 4-Bromofluorobenzene	11.14	95	93334	45.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.86%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	590	0.391	ug/l #	76
16) Acetone	2.44	43	6967	8.507	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	4074	2.350	ug/l	87
37) Ethyl Acetate	4.83	43	7868	3.800	ug/l #	89
44) Trichloroethene	7.20	130	16399	9.259	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102419\
Data File : VX013196.D
Acq On : 23 Oct 2019 18:55
Operator : JC/SP
Sample : K5451-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW0542

Quant Time: Oct 24 04:22:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

