

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032339.D
 Acq On : 24 Oct 2022 14:47
 Operator : JC/MD
 Sample : N5248-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-RB-20221018

Quant Time: Oct 25 03:27:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203153	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176569	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80046	46.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.400%
35) Dibromofluoromethane	5.391	113	65670	46.453	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.900%
50) Toluene-d8	8.653	98	236100	46.396	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.800%
62) 4-Bromofluorobenzene	11.079	95	80844	43.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.760%
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	2443	0.464	ug/l #	81
52) Toluene	8.720	92	10587	3.101	ug/l	98
67) Ethyl Benzene	10.195	91	1747	0.273	ug/l	96
68) m/p-Xylenes	10.305	106	2563	1.013	ug/l	95
69) o-Xylene	10.647	106	990	0.408	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	1171	0.239	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032339.D
Acq On : 24 Oct 2022 14:47
Operator : JC/MD
Sample : N5248-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-RB-20221018

Quant Time: Oct 25 03:27:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

