

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032343.D
 Acq On : 24 Oct 2022 16:20
 Operator : JC/MD
 Sample : N5237-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-GW-01

Quant Time: Oct 25 03:27:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	136314	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	215149	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84463	47.630	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.260%
35) Dibromofluoromethane	5.391	113	68426	45.704	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.400%
50) Toluene-d8	8.653	98	249792	46.350	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.700%
62) 4-Bromofluorobenzene	11.079	95	86098	43.628	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032343.D
Acq On : 24 Oct 2022 16:20
Operator : JC/MD
Sample : N5237-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-GW-01

Quant Time: Oct 25 03:27:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

