

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032344.D
 Acq On : 24 Oct 2022 16:43
 Operator : JC/MD
 Sample : N5236-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSP-GW-03

Quant Time: Oct 25 03:28:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203662	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81529	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	80610	47.148	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.300%
35) Dibromofluoromethane	5.391	113	66455	46.891	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.780%
50) Toluene-d8	8.653	98	237749	46.603	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.200%
62) 4-Bromofluorobenzene	11.079	95	81147	43.438	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.880%
Target Compounds						
16) Acetone	2.386	43	1143	2.060	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032344.D
Acq On : 24 Oct 2022 16:43
Operator : JC/MD
Sample : N5236-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSP-GW-03

Quant Time: Oct 25 03:28:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

