

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032347.D
 Acq On : 24 Oct 2022 17:53
 Operator : JC/MD
 Sample : N5260-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-10-21

Quant Time: Oct 25 03:58:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89979	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88602	48.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.640%
35) Dibromofluoromethane	5.385	113	70210	46.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.560%
50) Toluene-d8	8.653	98	253287	46.386	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.780%
62) 4-Bromofluorobenzene	11.079	95	89366	44.694	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.380%
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	14607	2.107	ug/l	95
68) m/p-Xylenes	10.305	106	7048	2.575	ug/l	99
78) n-propylbenzene	11.305	91	4200	0.571	ug/l	96
80) 1,3,5-Trimethylbenzene	11.457	105	2321	0.419	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	16062	2.979	ug/l	99
95) Naphthalene	13.774	128	6657	1.074	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032347.D
Acq On : 24 Oct 2022 17:53
Operator : JC/MD
Sample : N5260-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-10-21

Quant Time: Oct 25 03:58:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

